



Politecnico di Bari

Repository Istituzionale dei Prodotti della Ricerca del Politecnico di Bari

Experimental investigation of the mechanical performances of titanium cranial prostheses manufactured by super plastic forming and single-point incremental forming

This is a post print of the following article

Original Citation:

Experimental investigation of the mechanical performances of titanium cranial prostheses manufactured by super plastic forming and single-point incremental forming / Ambrogio, G., Palumbo, G., Sgambitterra, E., Guglielmi, P., Piccininni, A., de Napoli, L., Villa, T., Fragomeni, G.. - In: INTERNATIONAL JOURNAL, ADVANCED MANUFACTURING TECHNOLOGY. - ISSN 0268-3768. - STAMPA. - 98:5-8(2018), pp. 1489-1503. [10.1007/s00170-018-2338-6]

Availability:

This version is available at <http://hdl.handle.net/11589/136422> since: 2021-03-15

Published version

DOI:10.1007/s00170-018-2338-6

Publisher:

Terms of use:

(Article begins on next page)

Experimental investigation of the mechanical performances of Titanium cranial prostheses manufactured by Super Plastic Forming and Single Point Incremental Forming

G. Ambrogio^{1,*}, G. Palumbo², E. Sgambitterra¹, P. Guglielmi², A. Piccininni², L. De Napoli¹, T. Villa³, G. Fragomeni⁴

¹Dipartimento di Ingegneria Meccanica, Energetica e Gestionale, Università della Calabria

² Dipartimento di Meccanica, Matematica e Management, Politecnico di Bari,

³Dipartimento di Chimica, Materiali e Ingegneria Chimica "G. Natta", Politecnico di Milano

⁴Università di Catanzaro

Abstract

In the present work sheet forming processes, i.e. Super Plastic Forming and Single Point Incremental Forming, have been adopted for the manufacturing of custom prostheses, instead of subtractive and additive techniques, which result time and cost consuming for a single-piece production. As concerns the material, three different Titanium alloys were used: pure Titanium and two grades of the alloy Ti-6Al-4V (the standard one and the Extra Low Interstitial one).

Since no standard protocol exists to assess the mechanical performance of cranial implants, an experimental procedure has been designed and used in this work for producing polymethylmethacrylate supports on which the cranial prostheses were firmly connected and subjected to impact puncture tests (drop tests). An experimental campaign could thus be conducted to investigate the effect on the mechanical response of: (i) the titanium alloy, (ii) the initial blank thickness and (iii) the manufacturing process. Drop tests, carried out according to the proposed procedure, have shown no failure of the prostheses, neither in the area of the impact nor in the anchoring region and have revealed that, irrespective of the adopted manufacturing process, which does not alter the material, the amount of energy absorbed by the implants is always larger than 70%.

Keywords: *SPF; SPIF; pure titanium; Ti-6Al-4V; Ti-6Al-4V-ELI; Drop Test.*

*Corresponding author: g.ambrogio@unical.it

1. Introduction

Recent studies in manufacturing and bioengineering have confirmed the increasing demand for high-quality biomedical implants able to increase the life expectancy of patients and to avoid, at the same time, prolonged hospitalization [1]. Moreover, it is well known that custom implants, which are prostheses able to perfectly fit the damaged region of a specific patient, provide a mechanical connection good enough to avoid relative micro-motions with the surrounding bone region and to prevent the occurrence of infection sites [2]. In addition, custom implants are the gold standard solution when aesthetic compatibility is required, since they ensure very good aesthetic results [3,4]. Cranioplasty, for example, whose aim is to reconstruct structural or morphological cranial regions due to congenital or accidental causes, is a kind of neuro-reconstructive surgery and currently still a challenge for both surgeons and bioengineers. Actually, it is a clinical procedure not only for an anatomical reconstruction, but also for the neurological improvement of underlying physiology [5]. Matching the best material and the suitable manufacturing process to obtain a custom implant with higher biomechanical properties is not a trivial question.

Literature offers plenty of data regarding the adoption of several biomaterials, ranging from metals to bioceramics [6] and biopolymers [7], but titanium (Ti) and its alloys still remain the most widely adopted solutions since they possess a low value of the Young's modulus (thus assuring an homogeneous stress distribution between the implant and the surrounding bone [8,9]), excellent mechanical properties and moreover, as far as biocompatibility is concerned, the capability of promoting osseointegration [10,11].

Concerning the production process, stereolithography [12] and the CAD/CAM approach [13–15] are considered highly effective for manufacturing bioimplants since such processes facilitate, speed up and improve the quality of surgical procedures. On the other hand, especially when the geometry of the prosthesis is characterised by thin walls and extreme complexity (even undercuts), sheet metal forming processes, such as the Super Plastic Forming (SPF) and the Single Point Incremental Forming (SPIF) can be valid alternatives to the conventional subtractive ones [16]. In the SPF process, the characteristic of some fine-grained alloys to exhibit, under certain working conditions, very high ductility, resulting in extremely high deformation levels before fracture, is exploited. At its simplest a thin sheet is clamped into tools heated to a certain temperature level (typically of the order of half the absolute melting temperature) and deformed into the die within a controlled strain rate (e.g. range 10^{-5} to 10^{-1} s^{-1}) by means of an (inert) gas pressure in the order of

few MPa. SPF is characterized by low flow stress and high uniformity of plastic flow [18]; in addition, by managing the gas pressure profile, the thickness distribution of the component can be controlled [17]. SPIF process is characterised by a clamped sheet formed into a desired shape by the action imposed with a controlled movement of a hemispherical punch. According to that the deformation mechanism is mainly stretching and the obtained thickness is dependent by the complexity of the profile and the imposed manufacturing strategy [19]; SPIF is characterised by a flexibility higher than SPF and, at the same time, the acquired knowledge on this process have led to a more controlled accuracy [19–21].

It has to be pointed out that the overall quality of a prosthetic implant is based not only on its fitting and connection characteristics, but also on its structural properties. For this reason, it is fundamental to assess the load-bearing capacity of the implant. The structural requirements of medical devices can widely vary when the applied load conditions changes after the implantation. In some cases the magnitude of such loads can be very high (e.g. in a hip prosthesis it can reach five times the value of the body weight while in dental implants can reach almost 1kN) being then resolved in a stress state dependent on both the final dimension and the geometry of the implant. The high stresses that can arise in an implantable device, regardless of the material it is made of, can in turn promote different failure mechanisms which lead to replace the device. Among the different failure mechanisms related to mechanical issues, wear, static and fatigue failures are the most common. Examples can be found in the replacements of knee where wear is the most frequent cause of revision, due to very high contact stresses in the polyethylene insert [22]. As far as static failures are concerned, due to the high magnitude of contact pressures in the hip joint, some examples of collapses of the ceramic components of hip prostheses can be found [23], while fatigue failure still mostly affects the metallic stem [24] or neck of such a kind of prosthesis [25]. Other examples of failures due to demanding mechanical stimuli can be cited for different materials: fatigue failures of Nitinol peripheral stents still affect their survivorship [26] and expose the patient to the need for a new surgery. Most medical implants require very specific properties: an arterial graft, for instance, needs to offer flexibility, anisotropic behaviour and compliance which match that one of the adjacent vessel [27]; a balloon angioplasty catheter needs to be tractable over a guide wire and must be stiff enough to prevent kinking and flexible enough to navigate the vasculature; sutures need to provide high tensile strength and in case of resorbable sutures, this strength must decrease over time in a controlled manner; orthodontic wires require high elastic limit and low stiffness, because they provide the ability of applying lighter and more constant forces

during arch deactivation [28]. Cranial prostheses, instead, require low stiffness and high toughness in order to guarantee a good distribution of the stresses and to better absorb accidental impact [29]. For such reasons, the structural properties of the biomaterials and the way they are manufactured are extremely important for the long-term success of the implant. In addition, ad-hoc experimental setup and equipment are required to characterize the specific application of biomedical implants. For some of them, approved standards are available: for example, for evaluating the static and fatigue behaviour of lower limb orthopaedic prostheses [30] and devices for spinal surgery [31], or for determining the wear behaviour of components used in total joint prostheses [32], or for the measuring the corrosion fatigue of metallic implants [33].

On the other hand, when dealing with cranial implants, there is a lack of standards able to provide testing procedures for their mechanical characterization; such a lack is probably due to the fact that cranial implants do not bear loads during their use, but they can be only accidentally overloaded because of fortuitous falls or accidents in normal everyday activity. In this context, only few studies report procedures to assess the structural performance of cranial prostheses. For example, Tsouknidas et al. investigated the mechanical response of implants subjected to an additional load simulating the impact of a tennis ball at a reference speed of 30 m/s [34]; Garcia-Gonzalez et al. [35] simulated the impact conditions of an accidental fall (for instance from the bed or the bike) on polyether-ether-ketone and macro-porous hydroxyapatite implants.

In this work the mechanical performance of Ti-cranial prostheses manufactured by SPF and SPIF has been assessed by means of impact puncture tests carried out using a universal INSTRON drop test machine. For this aim, a proper setup has been designed to perform the impact experiments and three different Ti-alloys have been investigated: pure Titanium and two grades of the alloy Ti-6Al-4V (the standard one and the Extra Low Interstitial one). Since no standard actually exists for assessing the mechanical performance of cranial implants, a new experimental procedure has been designed and used in this work for producing polymethylmethacrylate (PMMA) supports on which the cranial prostheses were firmly connected and subjected to drop tests. An experimental campaign was thus conducted adopting two different load conditions in order to investigate the effect on the mechanical response of: (i) the titanium alloy, (ii) the blank initial thickness and (iii) the manufacturing process.

2. Material, experimental setup and procedure

2.1. The investigated case study

A schematic depiction of the entire process required for the geometrical definition of the prosthesis adopted, as case study, in the present work, is shown in *Figure 1*.

In particular, starting from a polymer model anatomically identical to the human skull, a defect in the apex frontal area was created (*Figure 1a*). The best prosthesis geometry, able to fill the bone defect, was defined by means of the damage acquisition (*Figure 1b*) and the subsequent reconstruction using mirroring techniques (*Figure 1c*). In addition, the prosthesis surface was a bit extended in order to ensure the optimal anchoring with the polymer model [36].




Figure 1. Case study.

Three Ti alloys, largely adopted for biomedical devices, were investigated in the present work: the commercial pure Ti (Ti-Gr2), the Ti-6Al-4V (Ti-Gr5) and the Extra Low Interstitial Ti-6Al-4V (Ti-Gr23). **Table 1 reports the chemical compositions and the Elastic moduli of the investigated alloys.**

Table 1. Modulus of Elasticity and chemical composition of the investigated Ti alloys

Alloy	E [GPa]	Al [%]	V [%]	Fe [%]	C [%]	N [%]	H [%]	O [%]	Ti [%]
Ti-Gr2	103	-	-	0.3 max	0.10 max	0.03 max	0.015 max	0.3 max	Bal.
Ti-Gr5	113	5.5 - 6.5	3.5 - 4.5	0.3 max	0.08 max	0.05 max	0.015 max	0.2 max	Bal.
Ti-Gr23	113	5.5 - 6.5	3.5 - 4.5	0.25 max	0.08 max	0.05 max	0.015 max	0.13 max	Bal.

Both the SPIF and SPF processes were adopted to manufacture prostheses made of Ti-Gr5 (1 and 1.5mm thick); on the contrary, prostheses made of Ti-Gr2 (1.5mm) and Ti-Gr23 (1mm) were manufactured only by SPIF and SPF, respectively. This choice is justified by

the opportunity of investigating both processes when the material properties change. According to that, Ti-Gr2 was adopted for testing the SPIF feasibility also in cold condition, whereas Ti-Gr23 was chosen for its superplastic behaviour, suitable for SPF operations. A prior investigation of all the adopted materials was made by using flat samples. In particular, preliminary tests were performed on Ti-Gr2 and Ti-Gr5 using flat samples 1.0 and 1.5 mm thick, and on Ti-Gr23, using flat samples 1.0 mm thick.

2.2. Hardness tests

Since blanks having the same Elastic modulus (two blanks cut from sheets made by the same alloy but having different thickness values or two blanks cut from sheets made by two different alloys having the same Elastic modulus) were investigated, the mechanical properties of the investigated Ti alloys in the as received conditions were assessed by Vickers microhardness tests. Tests were performed according to the UNI EN ISO 6507-1 Standard [37]. The HWMMT-X7 HIGHWOOD micro hardness tester was used and a load of 1000 g was chosen. Six replications were performed on each sample.

2.3. Experimental setup for prostheses manufacturing

SPF experiments were conducted using the 2500 kN prototypal electro-hydraulic press machine shown in *Figure 2a*. Forming tools were heated by means of electric cartridges managed by a PLC. As shown in *Figure 2b*, the upper tool was equipped with a metallic female insert having the geometry of the case study.




Figure 2. Experimental set-up for SPF.

Before each test, boron nitride was applied on both blank surfaces in order to simplify the extraction of the formed part. Temperatures were monitored during the process using K-type thermocouples positioned in both the upper and the lower tool. When the forming

temperature (850 °C) was reached [38], the blank was introduced between the tools and then clamped (a blankholder force of 490 kN was used for preventing any drawing and gas leakage). The blank was formed by the gas whose pressure was set according to an optimized profile obtained from numerical simulations carried out using a commercial finite element code (ABAQUS) and implementing the material superplastic behaviour [40,41]. In the present investigation, the numerical model was simply used for defining, for each of the adopted alloys (Ti-Gr23 1.0 mm thick, Ti-Gr5 1.0 mm thick and Ti-Gr5 1.5 mm thick), the pressure profiles able to keep the strain rate close to the optimal value (10^{-4} s^{-1}). The pressure vs time curves, obtained using the ABAQUS built-in subroutine and in Figure 3, were used as input parameters for the press machine.




Figure 3. Optimized pressure profile obtained from numerical simulation for the SPFed prostheses.

SPIF experiments were performed on a 3-axis CNC MAZAK milling machine. The adopted equipment is sketched in Figure 4.




Figure 4. Experimental set-up for SPIF.

As shown in *Figure 4a*, the standard SPIF clamping frame was equipped with an additional electrical heater which allowed to perform the process in hot-condition. The high temperature, in fact, is mandatory for obtaining sound results when processing Ti-Gr5. The original clamping frame was thus enriched with a 2kW electrical heater and insulated in order to reduce heat losses to the environment and to the working table (*Figure 4b*). In this way, by using the same equipment, the process was performed both at room temperature (without any heating source) for manufacturing Ti-Gr2 and in hot conditions for processing Ti-Gr5 [41]. To perform the hot SPIF in the proper way, two thermocouples positioned in the chamber and on the sheet were utilised to control the temperature. Preliminary experiments were conducted using a FLIR thermo-camera to check and properly calibrate the PID controller of the electrical heater: a temperature of 650°C was reached during the forming phase and properly kept constant during the manufacturing. An hemispherical HSS tool with a diameter of 12 mm was adopted. The sheet was deformed according to a tool trajectory previously defined by means of numerical simulations [16]. Regardless of the adopted temperature, all tests were performed using a constant tool pitch equal to 0.5 mm, a tool feed rate of 2.5 m/min and a tool rotational speed of 600 RPM.

After the manufacturing step, both SPFed and SPIFed parts were cut by means of a 3D laser in order to extract the final geometry from the formed blank; using the same fixture, the laser drilling of eight holes for the anchoring to the PMMA support was also performed. The laser cutting path and the position of the holes (approximately at an angular distance of 45° each other) are shown in *Figure 5*, while in *Figure 6* the set up adopted for the laser cutting operation is presented.

Preliminary trial cuts were performed on flat specimens in order to tune the laser parameters. Finally, the following ones (able to minimize burrs) were adopted: laser power=400 W, nitrogen pressure=12 bar, cutting speed=1 m/min, focal length=4 mm, stand-off distance=1.2 mm.




Figure 5. Cutting path and position of the holes.

Figure 6. Formed blank positioning for the 3D laser cutting of SPFed and SPIFed parts.

2.4. Drop tests

2.4.1. Preparation of the support and anchoring of the prosthesis

Drop tests were performed on SPFed and SPIFed prostheses connected on purposely made anchoring supports whose geometry was made identical to the upper portion of the defected skull. The procedure to use and the material the supports were made of, were chosen with the aim to produce a large number of them and, at the same time, to make the process robust and repeatable. The procedure illustrated in *Figure 7* shows the supports manufacturing using a synthetic resin (PMMA, also known as “bone cement”) which was poured into a SPFed formed blank used as mould.



Figure 7. Procedure of support manufacturing.

The PMMA was chosen as material for the supports realization because this biomaterial is commonly used to connect the prosthesis to the host bone since it has mechanical characteristics very similar to the human bone.

In order to mimic the defect inside the support, a proper shaped core, having approximately the dimension of the defect, was positioned in the cavity before pouring the PMMA. After the resin curing, the support was extracted and four bushes were inserted in the bottom surface so that it could be rigidly coupled with the standard clamping system of the drop test machine by means of four screws. The anchoring of the SPFed and SPIFed prostheses to the supports was made mimicking the surgical technique: as reported in *Figure 8*, four standard self-cutting 2.0 mm titanium cortical screws were used to connect the prosthesis to the PMMA support.




Figure 8. Anchoring of the prosthesis: a) PMMA support, b) anchored prosthesis.

2.4.2. Drop tests equipment

Drop tests were performed on the 3D laser cut prostheses using an Instron CEAST 9350 drop tower, which is able to perform impact experiments applying energies up to 1800 J by means of a spring assist. An instrumented striker equipped with a polished 20 ± 0.2 mm hemispherical nose impactor (mass: 2.295 kg) was employed in the present work. Due to the non-standard shape of the samples, the purposely made supports described in the section 2.4.1 were used to firmly connect the prostheses and thus to minimize vibrations during the test. In particular, the support on which the prosthesis was anchored, was connected to the clamping system shown in *Figure 9*, made of a plate having four columns able to carry an additional upper plate used to avoid vibrations during the impact.



Figure 9. Set up for drop tests.

All components were properly designed in order to guarantee high stiffness of the whole system. After the first impact of the specimen, a break mechanism automatically prevented a second strike. During the impact, the resistive force exerted by the specimen was measured by a 45 kN strain gauge load cell. Experiments were conducted adopting two different drop height (h) for each specimen type: 200 and 600 mm, corresponding to nominal impact energies (E_{max}) and velocities (v_i) of 4.5 J - 1.98 m/s and 13.5 J - 3.43 m/s, respectively. Such values can simulate a possible daily accident. At least three tests were performed for each load condition.

From the basic force–time information, important parameters such as deflection, $\delta(t)$, absorbed energy, $E(t)$, and velocity, $v(t)$ were calculated. The velocity versus time was evaluated by using Eq. (1) [42]. A positive velocity value indicates a downward motion.

$$v(t) = v_i + gt - \int_0^t \frac{1}{m} F(t) dt \quad (1)$$

where t is the time; $v(t)$ and v_i are the impactor velocities at time t and $t=0$, respectively; $F(t)$ is the measured impactor contact force at time t ; m is the mass of impactor. The deflection $\delta(t)$ versus time was calculated through Eq. (2) [42]:

$$\delta(t) = \delta_i + v_i t + \frac{gt^2}{2} - \int_0^t \int_0^t \frac{1}{m} F(t) dt \quad (2)$$

where $\delta(t)$ and δ_i are the impactor displacements from the reference location at time t and $t=0$, respectively. The absorbed energy $E(t)$ versus time was calculated through Eq. (3) [42]:

$$E(t) = \frac{m(v_i^2 - v(t)^2)}{2} + mg\delta \quad (3)$$

where $E(t)$ is the absorbed energy at time t .

2.4.3. Output of the drop tests

Figure 10 shows the typical responses obtained from an instrumented impact test. In particular, Figure 10a shows both the force and the energy response according to time, while Figure 10b shows the acquired force according to the sample deflection.




Figure 10. Responses obtained from instrumented impact test: a) force and energy vs time, b) force vs deflection.

In the figures above, all the investigated parameters are also shown: F_{max} (the maximum force), E_{max} (the maximum amount of energy that is absorbed by the specimen), E_r (the rebound energy, i.e. the energy returned by the sample to the punch), E_a (the total energy absorbed by the sample), δ_{max} (the maximum deflection), δ_{res} (the residual deflection) and δ_{rec} (the recovered deflection). In order to compare the behaviour of different materials, even characterised by a different thickness, the absorbed energy ratio (i.e. the absorbed energy over the maximum applied energy, E_a/E_{max}) was additionally considered in the present work.

3. Results and Discussion

3.1. Assessment of mechanical characteristics of the as received sheets

Hardness results concerning samples extracted from the sheets of the investigated Ti-alloys in the as-received condition are plotted in *Figure 11* as mean values.




Figure 11. Micro-hardness average values of the investigated Ti-alloy in the as received condition.

The hardness levels revealed that the thickness always affects the mechanical behaviour, even in case of the same alloy. For example, Ti-Gr5 1.0 mm thick has a hardness level about 10% greater than Ti-Gr5 1.5 mm thick. The opposite can be noted when comparing the hardness of the two samples in Ti-Gr2: the 1mm thick has a hardness level about 10% lower than the 1.5mm. Such differences appear to be a consequence of the production process the sheet was subjected to, thus revealing the necessity of assessing the material behaviour before the investigation on the formed blanks.

3.2. Drop tests on flat samples

Preliminary tests, using the drop tower described in the section 2.4.2, were performed on flat samples in order to evaluate the impact response of the investigated materials in the as received condition. In particular, disk-shaped samples with a diameter of 60 mm were cut from the sheets and clamped on a support ring with an internal diameter of 40 ± 2 mm. The same testing parameters adopted to analyse the prostheses (see section 2.4) were used.

Table 2 resumes the obtained results (average data) in terms of: (i) energy, (ii) deflection and (iii) maximum force acquired during the tests; furthermore, the absorbed energy ratio (E_a/E_{max}) is also shown.

Table 2. Resume of results obtained from impact tests on flat samples.

Specimen	Falling height [mm]	E_{max} [J]	E_r [J]	E_a [J]	F_{max} [N]	δ_{max} [mm]	δ_{res} [mm]	δ_{rec} [mm]	E_a/E_{max}
Gr2-1mm	200	4.50	0.50	4.00	3154.79	2.96	2.49	0.47	0.89
	600	13.50	0.80	12.70	6327.48	4.52	4.2	0.32	0.94
Gr2-1.5mm	200	4.50	0.60	3.90	3666.86	1.97	1.56	0.41	0.86
	600	13.50	0.95	12.55	6946.98	3.43	3.04	0.39	0.93
Gr5-1mm	200	4.50	1.39	3.11	3749.22	2.53	1.42	1.11	0.69
	600	13.50	2.78	10.72	8117.94	3.82	2.65	1.17	0.79
Gr5-1.5mm	200	4.50	1.27	3.23	4092.99	1.91	1.07	0.85	0.72
	600	13.50	2.07	11.43	7892.34	3.24	2.20	1.04	0.85
Gr23-1mm	200	4.50	1.10	3.40	3720.57	2.66	1.79	0.87	0.75
	600	13.50	2.25	11.25	7856.53	4.06	3.10	0.97	0.83

The graph in Figure 12 show the effect of the thickness on the maximum deflection when using samples made of different alloys (Gr2 and Gr5) and setting different load conditions.




Figure 13. Effect of the thickness on the maximum deflection.




Figure 14. Effect of hardness on the ratio E_a/E_{max} .

It is possible to note that when increasing the thickness, irrespective to the material strength revealed by hardness measurements (*Figure 11*), the maximum deflection level was reduced (it is simply related to the Young's modulus and to the thickness) as a consequence of the stiffness increase of the sample.

In *Figure 15*, on the contrary, the effect of hardness on the ratio E_a/E_{max} is shown for both Gr2 and Gr5 and for the two load conditions: it is worthy of notice that such a parameter inversely affects the amount of energy stored by the sample, irrespective of the thickness value. In fact, when the value of the strength is higher one could expect a lower capability in absorbing energy. However, it is important to observe that ratio E_a/E_{max} ratio is always higher than about 70%. In fact, the smallest value was obtained in the test on the Gr5 1 mm thick (69%): it can be interpreted as a consequence of the highest measured hardness compared to the other Ti-alloys.

The graphs in *Figure* show the comparison among the three investigated Ti-alloys (thickness kept constant to 1.0 mm) in terms of maximum deflection (*Figure a*) and absorbed energy ratio (*Figure b*).




Figure 16. Comparison between the three investigated Ti-alloys (thickness of 1.0 mm).

It is possible to note that: (i) the investigated parameters are directly affected by the load condition; (ii) their values increase according to the height of fall. Furthermore, results show that Ti-Gr2 exhibits the highest deformation, as highlighted by the maximum deflection value reported in *Figure a* for both the load conditions, but also the highest capability in absorbing energy compared to the other alloys (*Figure b*). Such results are in good agreement with the mechanical response assessed by micro-hardness tests (*Figure 11*), which revealed that Ti-Gr2 possess the lowest hardness.

3.3. Drop Tests on cranial prostheses

An extensive experimental campaign was conducted on the prostheses manufactured using the investigated sheet forming processes. In particular, each cranial implant was tested using the setup and the testing conditions detailed in section 2.4. Three replications for each test were performed in order to reduce the sampling error. According to *Table 3*, six different configurations were investigated and a total amount of 36 cranial prostheses were tested.

Table 3. Summary of the performed drop test

	Process	SPF		SPIF	
	Height of fall	200mm	600mm	200mm	600mm
Material Type	Ti-Gr2, 1.5mm	-	-	Done	Done
	Ti-Gr5, 1mm	Done	Done	Done	Done
	Ti-Gr5, 1.5mm	Done	Done	Done	Done
	Ti-Gr23, 1mm	Done	Done	-	-

A typical response obtained from the impact test is reported in *Figure 17*: in particular, *Figure 17a* shows the acquired force during the test conducted on the prosthesis made of Ti-Gr23 1.0 mm thick subjected to an impact energy of 13.5 J, while in *Figure 17b* the force and the deflection acquired in the same test have been combined; furthermore, in both figures analogous results obtained from the test conducted on a flat sample (same alloy and thickness) have been plotted.




Figure 17. Responses from instrumented impact test performed on Ti-Gr23, 1mm thick: a) force vs time, b) force vs deflection.

Figure 17 highlights that both samples (the flat and the formed one) exhibit a similar response (increasing trend) only up to a certain force level; after that, bending instability occurred on the prosthesis, due to its specific shape, thus determining an evident change in the slope of the curves.

In Table 4 a summary of the main results obtained from the drop tests, in terms of average values, are presented.

Table 4. Summary of the results obtained from impact tests on cranial prostheses

Specimen	Falling height [mm]	E_{max} [J]	E_R [J]	E_a [J]	F_{max} [N]	δ_{max} [mm]	δ_{res} [mm]	δ_{rec} [mm]	E_a/E_{max}
SPIF_Gr2-1.5mm	200	4.50	0.60	3.90	3058.10	1.90	1.50	0.39	0.87
	600	13.50	1.15	12.35	4218.32	4.42	3.87	0.55	0.91
SPIF_Gr5-1mm	200	4.50	1.16	3.34	2159.29	3.54	2.30	1.24	0.75
	600	13.50	3.00	10.50	3303.40	6.47	4.08	2.39	0.78
SPIF_Gr5-1.5mm	200	4.50	1.00	3.50	3552.27	1.98	1.33	0.65	0.78
	600	13.50	2.18	11.32	4712.49	4.13	3.13	1.00	0.84
SPF_Gr5-1mm	200	4.50	1.85	2.65	1400.14	4.97	2.17	2.80	0.59
	600	13.50	4.20	9.30	2461.88	9.09	3.98	5.11	0.69
SPF_Gr5-1.5mm	200	4.50	1.19	3.31	2503.06	2.85	1.67	1.18	0.74
	600	13.50	3.00	10.50	3484.23	6.13	3.84	2.29	0.78
SPF_Gr23-1mm	200	4.50	1.60	2.90	1514.73	4.35	2.00	2.35	0.65
	600	13.50	3.05	10.45	2399.22	9.69	4.93	4.76	0.76

It is worthy of notice that all the investigated implants safely absorbed the applied energy without any crack insurgence: no fracture occurred neither in the region of the impact nor in the anchoring region where titanium cortical screws, connecting the prosthesis to the PMMA support, were positioned.

3.3.1. *Effect of the Ti-alloy*

The graphs in *Figure 18* show a direct comparison of the drop test results in terms of maximum deflection, maximum load and ratio E_a/E_{max} concerning the SPFed prostheses in Ti-Gr5 and Ti-Gr23 (both with an initial thickness of 1.0mm) when subjected to the two load conditions.




Figure 18. Comparison between SPFed prostheses in Ti-Gr23 and Ti-Gr5 (1.0 mm thick).

Figure 19 shows the same comparison but concerning the SPIFed prostheses in Ti-Gr5 and Ti-Gr2 (both with an initial thickness of 1.5mm).




Figure 19. Comparison between SPIFed prostheses in Ti-Gr5 and Ti-Gr2 (1.5 mm thick).

Focusing the attention on the highest load conditions, results in *Figure 18* and *Figure 19* show that the implants in Ti-Gr5 (manufactured by both SPF and SPIF) exhibited the lowest capability to absorb energy and, consequently, the lowest deformation. This result is

coherent with the **hardness measurements** presented in *Figure 11*, which show that Ti-Gr5 1.0 mm thick possesses a hardness level higher than Ti-Gr23 1.0 mm thick; in a similar way, Ti-Gr5 1.5 mm has a hardness level higher than Ti-Gr2 1.5 mm thick.

It is also important to note that the preliminary hardness measurements can only be used to give an interpretation of the results since they do not to give a comprehensive evaluation of the implant mechanical response. In fact, even though the hardness measurements revealed a marked difference between Ti-Gr5 1.5 mm thick and Ti-Gr2 1.5 mm thick, quite similar results were obtained for Ti-Gr5 1.0 mm thick and Ti-Gr23 1.0 mm thick. Moreover, data reported in *Figure 18* and *Figure 19* reveal, for both falling heights, a difference in terms of absorbed energy ratio close to 9.5%.

3.3.2. Effect of the initial blank thickness

The graphs in *Figure 20* show a direct comparison of the drop test results in terms of maximum deflection, maximum load and ratios E_a/E_{max} , concerning the SPFed prostheses in Ti-Gr5 with a thickness of 1.0 mm and 1.5 mm subjected to the two load conditions.

Results in *Figure 20* and in *Figure 21* reveal that for the prostheses made by alloy TiGr5, as observed during the impact tests performed on flat samples, the thicker the blank the lower the deflection and the higher the absorbed energy ratio. Such results reveal the possibility of assessing the effect of the initial blank thickness on drop test results simply using flat samples. However, the exact values of the considered test parameters strongly depend on the sample profile and, consequently, on the manufacturing process.




Figure 20. Comparison between prostheses in Ti-Gr5 with thickness 1.0 mm and 1.5 mm produced by SPF.

Figure 21 shows the same comparison but for the SPIFed prostheses in Ti-Gr5.




Figure 21. Comparison between prostheses in Ti-Gr5 with thickness 1.0 mm and 1.5 mm produced by SPIF.

3.3.3. Effect of the manufacturing process

In Figure 22 results from drop tests conducted on both SPFed and SPIFed prostheses in Ti-Gr5 1.5 mm thick have been compared. Figure 22a shows the trend of the maximum force and of the maximum deflection after the impact for both load conditions.

It can be noted that, in both testing cases, the implant produced by SPIF reached a maximum force about 40% larger than the one of the SPFed prosthesis (see continuous lines connecting full squares and dots respectively); in addition, SPIFed implants showed a maximum deflection δ_{max} smaller than the SPFed prosthesis one (when setting the falling height to 200mm, δ_{max} for the SPIFed prosthesis was about 40% lower than the SPFed one, whereas when setting the falling height to 600mm, δ_{max} for the SPIFed prosthesis was 48% lower than the SPFed one).




Figure 22. Comparison between prostheses in Ti-Gr5 (1.5 mm thick) produced by SPF and SPIF.

These results reveal that the implant created by SPIF can be considered more rigid than the one produced by SPF, but it is worthy of notice that, as shown in *Figure 22b*, the absorbed energy ratio exhibited by the SPIFed prosthesis are, for both load conditions, higher than the ones reached by the SPFed prosthesis.

Such results could be related to the material alteration determined by the manufacturing processes. To evaluate this effect, the thickness distributions along the longitudinal section of the samples manufactured both by SPF and by SPIF have been measured and results have been plotted in *Figure 21*. It is important to point out that the shown thickness distributions have been replicated (3 measurements) but also supported by numerical simulations [17].




Figure 23. Thickness distribution along the longitudinal section of the sample manufactured by (a) SPF and (b) SPIF (initial sheet thickness 1.5mm).

It can be noted that the investigated processes generated different thickness distributions and, as a consequence, a different mean value of the thickness. In particular, the SPIFed prosthesis, which is characterised by a mean thickness greater than the SPFed one, presents, in accordance with the results on flat samples, a lower maximum deflection and a higher maximum force.

On the contrary, the central region (where the impact load was applied) of the SPFed prostheses appears thinned (from 1.5 to almost 1.0 mm). Thus, in accordance with drop test data of flat samples, the thickness reduction, due to the manufacturing process, determined higher mechanical strength and, as a consequence, a lower level of the energy absorbed by the SPFed prosthesis (i.e. its capability to deform).

Finally, it can be concluded that, irrespective of both the production process and the initial thickness, the amount of energy absorbed by the implant when setting the largest height of fall (600 mm) was always larger than about 70% (the minimum value is 69% and concerns the prosthesis made of Ti-Gr5 1.0 mm thick fabricated by SPF); in addition, no fracture

occurred in any implant after the impact.

4. Conclusions

The medical device market requests of new and more performing devices for the treatment of different pathologies are increasing every day. Consequently, the necessity of investigating the mechanical reliability of implantable devices in the preclinical phase is becoming a more and more important activity, as to avoid that new devices are put into market without having demonstrated their safety. This lead to a continuous development and validation of new preclinical protocols that are often missing, especially for new devices.

In this manuscript, an additional brick of knowledge was introduced in the above mentioned direction: an innovative procedure was carefully designed for investigating the mechanical behaviour of titanium custom-made prostheses, manufactured by two innovative but well assessed technologies (Super Plastic Forming and Single Point Incremental Forming).

More deeply, the functional compatibility and the mechanical strength of the cranial implants were assessed by means of impact puncture tests. A specific set-up, using a purposely made PMMA anchoring support of geometry identical to the upper portion of the defected skull, was designed, fabricated and used since actually no standard exists in the literature. The tests allowed to effectively determine the mechanical response of the implant, since no prosthesis failed in the anchoring region where self-cutting titanium cortical screws were used for fixing purposes. The test **could thus be** used for investigating how implants made of different Titanium alloys (Ti-Gr2, Ti-Gr5 and Ti-Gr23) behave when subjected to an impact load, even if they are characterized by different thickness values and manufactured using different processes.

Results from drop tests revealed that, irrespective to the material strength (revealed by preliminary hardness measurements), the maximum deflection decreased when increasing the thickness. On the contrary, irrespective of the thickness value, the hardness inversely affected the ratio E_a/E_{max} (i.e. the amount of energy which is stored by the sample).

Drop tests also allowed to demonstrate that both the adopted manufacturing process did not alter significantly the material characteristics: even if the two forming processes produced two implants having different characteristics in terms of thickness distributions, the amount of absorbed energy, **in the most critical operating condition (i.e. when setting**

the height of fall to 600 mm), was always greater than about 70% (up to 85% for the prosthesis produced by Single Point Incremental Forming).

The proposed procedure for testing cranial prostheses can be thus considered to play a fundamental role for evaluating the implant mechanical response, also having the possibility to be extended to any implant geometry or material. In fact, the results from drop tests revealed to be in good agreement with the original mechanical behaviour of the investigated alloys: but the preliminary hardness measurements can simply give an interpretation of drop test results, being the comprehensive evaluation of the implant mechanical response demanded to drop tests due to the key role played by the sample profile and, most of all, by the thickness distribution. It would be thus interesting to investigate, in future works, the possibility to obtain a desired prosthesis performance (for example in terms of the parameter $\delta_{\max}$ or $E_a/E_{\max}$) preferably adopting a numerical approach. In fact, not only the real shape of the formed component, but also its thickness distribution can be imported in a numerical model for the dynamic explicit impact simulations. The desired level of the parameter $\delta_{\max}$ or $E_a/E_{\max}$ should be thus obtained by properly changing the manufacturing parameters and the process conditions in order to get the most suitable thickness distribution of the prosthesis to be tested.

Acknowledgements

The activities in this work were funded by the Italian Ministry of Education, Universities and Research Government through the PRIN Project 2012 “Biomedical Titanium alloy prostheses manufacturing by means of Superplastic and Incremental Forming processes” (project acronym: BIOFORMING).

References

- [1] L.A. Pruitt, A.M. Chakravartula, *Mechanics of Biomaterials: Fundamental Principles for Implant Design*, Cambridge University Press, Cambridge, 2011. doi:DOI: 10.1017/CBO9780511977923.
- [2] R.M. Wazen, J.A. Currey, H. Guo, J.B. Brunski, J.A. Helms, A. Nanci, Micromotion-induced strain fields influence early stages of repair at bone-implant interfaces, *Acta Biomater.* 9 (2013) 6663–6674. doi:10.1016/j.actbio.2013.01.014.
- [3] J. Castelan, L. Schaeffer, A. Daleffe, D. Fritzen, V. Salvato, F.P. Da Silva, Manufacture of custom-made cranial implants from DICOM?? images using 3D printing, CAD/CAM technology and incremental sheet forming, *Rev. Bras. Eng. Biomed.* 30 (2014) 265–

273. doi:10.1590/rbeb.2014.024.

- [4] H.R. Cho, T.S. Roh, K.W. Shim, Y.O. Kim, D.H. Lew, I.S. Yun, Skull reconstruction with custom made three-dimensional titanium implant, *Arch. Craniofacial Surg.* 16 (2015) 11–16. doi:10.7181/acfs.2015.16.1.11.
- [5] H. Eufinger, M. Wehmöller, a Harders, L. Heuser, Prefabricated prostheses for the reconstruction of skull defects., *Int. J. Oral Maxillofac. Surg.* 24 (1995) 104–110. doi:10.1016/S0901-5027(05)80870-7.
- [6] G. Staffa, A. Barbanera, A. Faiola, M. Fricia, P. Limoni, R. Mottaran, B. Zanotti, R. Stefani, Custom made bioceramic implants in complex and large cranial reconstruction: A two-year follow-up, *J. Cranio-Maxillofacial Surg.* 40 (2012) e65-70. doi:10.1016/j.jcms.2011.04.014.
- [7] S.-C. Lee, C.-T. Wu, S.-T. Lee, P.-J. Chen, Cranioplasty using polymethyl methacrylate prostheses, *J. Clin. Neurosci.* 16 (2009) 56–63. doi:10.1016/j.jocn.2008.04.001.
- [8] H.J. Rack, J.I. Qazi, Titanium alloys for biomedical applications, *Mater. Sci. Eng. C.* 26 (2006) 1269–1277. doi:10.1016/j.msec.2005.08.032.
- [9] M. Niinomi, Mechanical biocompatibilities of titanium alloys for biomedical applications, *J. Mech. Behav. Biomed. Mater.* 1 (2008) 30–42. doi:10.1016/j.jmbbm.2007.07.001.
- [10] M. Abdel-Hady Gepreel, M. Niinomi, Biocompatibility of Ti-alloys for long-term implantation, *J. Mech. Behav. Biomed. Mater.* 20 (2013) 407–415. doi:10.1016/j.jmbbm.2012.11.014.
- [11] D.R. Calderoni, R. Gilioli, A.L.J. Munhoz, R. Maciel Filho, C.A.D.C. Zavaglia, C.S. Lambert, E.S.N. Lopes, I.F.C. Toro, P. Kharmandayan, Paired evaluation of calvarial reconstruction with prototyped titanium implants with and without ceramic coating., *Acta Cir. Bras.* 29 (2014) 579–587. doi:10.1590/S0102-8650201400150005.
- [12] F.P.W. Melchels, J. Feijen, D.W. Grijpma, A review on stereolithography and its applications in biomedical engineering., *Biomaterials.* 31 (2010) 6121–30. doi:10.1016/j.biomaterials.2010.04.050.
- [13] W.-M. Xi, A.-M. Wang, Q. Wu, C.-H. Liu, J.-F. Zhu, F.-F. Xia, An integrated CAD/CAM/robotic milling method for custom cementless femoral prostheses, *Med. Eng. Phys.* 37 (2015) 911–915. doi:10.1016/j.medengphy.2015.06.005.
- [14] P. Kasprzak, G. Tomaszewski, G. Wróbel-Wiśniewska, M. Zawirski, Polypropylene-polyester cranial prostheses prepared with CAD/CAM technology. Report of first 15 cases., *Clin. Neurol. Neurosurg.* 113 (2011) 311–315. doi:10.1016/j.clineuro.2010.12.010.
- [15] J.-J. Chen, W. Liu, M.-Z. Li, C.-T. Wang, Digital manufacture of titanium prosthesis for cranioplasty, *Int. J. Adv. Manuf. Technol.* 27 (2006) 1148–1152. doi:10.1007/s00170-004-2309-y.
- [16] A. Piccininni, F. Gagliardi, P. Guglielmi, L. De Napoli, G. Ambrogio, D. Sorgente, Biomedical Titanium alloy prostheses manufacturing by means of Superplastic and Incremental Forming processes, 15007 (2016). doi:10.1051/mateconf/201.

- [17] D. Sorgente, G. Palumbo, A. Piccininni, P. Guglielmi, S.A. Aksenov, Investigation on the thickness distribution of highly customized titanium biomedical implants manufactured by superplastic forming, *CIRP J. Manuf. Sci. Technol.* 20 (2018). doi:10.1016/j.cirpj.2017.09.004.
- [18] M. Jackson, Superplastic forming and diffusion bonding of titanium alloys, in: *Superplast. Form. Adv. Met. Mater.*, Elsevier, 2011: pp. 227–246. doi:10.1533/9780857092779.3.227.
- [19] T. Madeira, C.M.A. Silva, M.B. Silva, P.A.F. Martins, Failure in single point incremental forming, *Int. J. Adv. Manuf. Technol.* 80 (2015) 1471–1479. doi:10.1007/s00170-014-6381-7.
- [20] S. Gatea, H. Ou, G. McCartney, Review on the influence of process parameters in incremental sheet forming, *Int. J. Adv. Manuf. Technol.* 87 (2016) 479–499. doi:10.1007/s00170-016-8426-6.
- [21] D. Afonso, R. Alves de Sousa, R. Torcato, Integration of design rules and process modelling within SPIF technology-a review on the industrial dissemination of single point incremental forming, *Int. J. Adv. Manuf. Technol.* 94 (2018) 4387–4399. doi:10.1007/s00170-017-1130-3.
- [22] M.M. Landy, P.S. Walker, Wear of ultra-high-molecular-weight polyethylene components of 90 retrieved knee prostheses, *J. Arthroplasty.* 3 (1988) S73–S85. doi:10.1016/S0883-5403(88)80013-5.
- [23] A. Toni, F. Giardina, G. Guerra, A. Sudanese, M. Montalti, S. Stea, B. Bordini, 3rd generation alumina-on-alumina in modular hip prosthesis: 13 to 18 years follow-up results, *Hip Int.* 27 (2017) 8–13. doi:10.5301/hipint.5000429.
- [24] M. Junnila, I. Laaksonen, A. Eskelinen, P. Pulkkinen, L. Ivar Havelin, O. Furnes, A. Marie Fenstad, A.B. Pedersen, S. Overgaard, J. Kärrholm, G. Garellick, H. Malchau, K.T. Mäkelä, Implant survival of the most common cemented total hip devices from the Nordic Arthroplasty Register Association database, *Acta Orthop.* 87 (2016) 546–553. doi:10.1080/17453674.2016.1222804.
- [25] A. Merini, A. Viste, R. Desmarchelier, M.H. Fessy, Cementless Corailtm femoral stems with laser neck etching: Long-term survival, rupture rate and risk factors in 295 stems, *Orthop. Traumatol. Surg. Res.* 102 (2016) 71–76. doi:10.1016/j.otsr.2015.10.009.
- [26] A.R. Pelton, V. Schroeder, M.R. Mitchell, X.Y. Gong, M. Barney, S.W. Robertson, Fatigue and durability of Nitinol stents, *J. Mech. Behav. Biomed. Mater.* 1 (2008) 153–164. doi:10.1016/j.jmbbm.2007.08.001.
- [27] M. Hasegawa, T. Azuma, Mechanical properties of synthetic arterial grafts, *J. Biomech.* 12 (1979) 509–517. doi:10.1016/0021-9290(79)90039-3.
- [28] S. Kapila, R. Sachdeva, Mechanical properties and clinical applications of orthodontic wires., *Am. J. Orthod. Dentofac. Orthop.* 96 (1989) 100–9. doi:http://dx.doi.org/10.1016/0889-5406(89)90251-5.
- [29] B. Zanotti, N. Zingaretti, A. Verlicchi, M. Robiony, A. Alfieri, P.C. Parodi, Cranioplasty, *J. Craniofac. Surg.* 27 (2016) 2061–2072. doi:10.1097/SCS.00000000000003025.

- [30] ISO 7206-4:2010, Implants for surgery -- Partial and total hip joint prostheses -- Part 4: Determination of endurance properties and performance of stemmed femoral components., (n.d.).
- [31] ISO 12189:2008, Implants for surgery -- Mechanical testing of implantable spinal devices -- Fatigue test method for spinal implant assemblies using an anterior support, (n.d.).
- [32] ASTM F732 - 00(2011), Standard Test Method for Wear Testing of Polymeric Materials Used in Total Joint Prostheses, (n.d.).
- [33] ASTM F1801 - 97(2014), Standard Practice for Corrosion Fatigue Testing of Metallic Implant Materials, (n.d.).
- [34] A. Tsouknidas, S. Maropoulos, S. Savvakis, N. Michailidis, FEM assisted evaluation of PMMA and Ti6Al4V as materials for cranioplasty resulting mechanical behaviour and the neurocranial protection, *Biomed. Mater. Eng.* 21 (2011) 139–147. doi:10.3233/BME-2011-0663.
- [35] D. Garcia-Gonzalez, J. Jayamohan, S.N. Sotiropoulos, S.H. Yoon, J. Cook, C.R. Siviour, A. Arias, A. Jérusalem, On the mechanical behaviour of PEEK and HA cranial implants under impact loading, *J. Mech. Behav. Biomed. Mater.* 69 (2017) 342–354. doi:10.1016/j.jmbbm.2017.01.012.
- [36] C. Ciancio, M.V. Caruso, G. Fragomeni, G. Ambrogio, Integrated thermomechanical model for forming glass containers, *MATEC Web Conf.* 80 (2016) 16010-1-16010-6. doi:10.1051/mateconf/201.
- [37] UNI EN ISO 6507, Metallic materials - Vickers hardness test - Part 1: Test method, (n.d.).
- [38] H.-S. Lee, J.-H. Yoon, C.H. Park, Y.G. Ko, D.H. Shin, C.S. Lee, A study on diffusion bonding of superplastic Ti-6Al-4V ELI grade, *J. Mater. Process. Technol.* 187–188 (2007) 526–529. doi:10.1016/j.jmatprotec.2006.11.215.
- [39] J. Bonet, R.D. Wood, R. Collins, Pressure-control algorithms for the numerical simulation of superplastic forming, *Int. J. Mech. Sci.* 36 (1994) 297–309. doi:10.1016/0020-7403(94)90036-1.
- [40] D. Sorgente, G. Palumbo, A. Piccininni, P. Guglielmi, L. Tricarico, Modelling the superplastic behaviour of the Ti6Al4V-ELI by means of a numerical/experimental approach, *Int. J. Adv. Manuf. Technol.* 90 (2017) 1–10. doi:10.1007/s00170-016-9235-7.
- [41] A. Giuseppina, C. Claudio, F. Luigino, G. Francesco, Theoretical model for temperature prediction in Incremental Sheet Forming – Experimental validation, *Int. J. Mech. Sci.* 108–109 (2016) 39–48. doi:10.1016/j.ijmecsci.2016.01.030.
- [42] ASTM Standard D. 7136; 2005, Standard test method for measuring the damage resistance of a fiberreinforced polymer matrix composite to a drop-weight impact event, (n.d.).